



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Heat Treating Services Unlimited, Inc.
222 LaDean Court, Suite G
Simpsonville, SC 29681

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

R. Douglas Leonard Jr., VP, PILR SBU

Expiry Date: 03 August 2024

Certificate Number: L2138



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Heat Treating Services Unlimited, Inc.

222 LaDean Court, Suite G
Simpsonville, SC 29681
Neil Revis 864-289-0644

CALIBRATION

Valid to: **August 3, 2024**

Certificate Number: **L2138**

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Source/Measure	(0 to 4) mA (4 to 20) mA	1.5 μ A 3.8 μ A	Martel 3001 Calibrator with Agilent 3458A 8.5 Digit Multimeter
DC Voltage – Source/Measure	(0 to 110) mV (0.11 to 1.1) V (1.1 to 11) V	21 μ V 6.1 mV 0.72 mV	
Resistance Simulation of RTD Indicators – Source/Measure	Pt 385 100 Ω (0 to 400) $^{\circ}$ C (400 to 800) $^{\circ}$ C Pt 385 1 000 Ω (0 to 195) $^{\circ}$ C	0.11 $^{\circ}$ C 0.062 $^{\circ}$ C 0.22 $^{\circ}$ C	
Electrical Simulation of Thermocouple Indicators – Source/Measure ¹	Type B (100 to 593) $^{\circ}$ C (593 to 849) $^{\circ}$ C (849 to 1 301) $^{\circ}$ C Type E (-200 to 0) $^{\circ}$ C (0 to 982) $^{\circ}$ C Type J (-100 to 800) $^{\circ}$ C (800 to 1 200) $^{\circ}$ C Type K (-100 to 400) $^{\circ}$ C (400 to 1 372) $^{\circ}$ C Type N (-100 to 900) $^{\circ}$ C (900 to 1 300) $^{\circ}$ C	0.46 $^{\circ}$ C 0.44 $^{\circ}$ C 0.59 $^{\circ}$ C 0.21 $^{\circ}$ C 0.31 $^{\circ}$ C 0.29 $^{\circ}$ C 0.32 $^{\circ}$ C 0.28 $^{\circ}$ C 0.3 $^{\circ}$ C 0.31 $^{\circ}$ C 0.3 $^{\circ}$ C	Thermocouple Half Junction with Agilent 3458A 8.5 Digit Multimeter and Ice point

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicators – Source/Measure ¹	Type R (-20 to 0) °C	0.42 °C	Thermocouple Half Junction with Agilent 3458A 8.5 Digit Multimeter and Ice point
	(0 to 100) °C	0.4 °C	
	(100 to 1 750) °C	0.41 °C	
	Type S (0 to 200) °C	0.42 °C	
	(200 to 1 400) °C	0.4 °C	
	(1 400 to 1 752) °C	0.41 °C	
Type T (-200 to 0) °C		0.29 °C	
	(0 to 400) °C	0.32 °C	

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure – Hydraulic ¹	(0 to 10 000) psig	5 psi	Comparison to Fluke 700G Pressure Gage

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Uniformity Survey of Furnaces & Ovens ¹	(0 to 1 093) °C (1 093 to 1 250) °C	1.1 °C 2.7 °C	In accordance with AMS 2750E using a Datalogger and Type K thermocouples
Temperature System Accuracy Tests ¹	Type K (0 to 1 093) °C	1.2 °C	Thermocouple Calibrator with reference TC wire in accordance with AMS 2750E
	(1 093 to 1 250) °C	2.2 °C	
	Type N (0 to 1 093) °C	1.3 °C	
	(1 093 to 1 250) °C	2.1 °C	

Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Timers ¹	Up to 1 min (1 to 30) min (30 to 60) min	2.9 s 4.6 s 4.6 s	Reference Stopwatch

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. This scope is formatted as part of a single document including Certificate of Accreditation No. L2138.



R. Douglas Leonard Jr., VP, PILR SBU

